

Anemometer Calibration for Wind Energy Applications

Rachael Coquilla
Otech Engineering, Inc., Davis, CA

Measurement Science Conference
March 2010

Topics for Discussion

- **Wind Speed Measurements in Wind Energy**
 - **Importance of wind speed measurements**
- **Anemometer Calibration Protocol**
 - **Calibration test standards**
- **Anemometer Calibration Uncertainty**
 - **IEC 61400-12-1 Uncertainty Analysis**
 - **Expanded Uncertainty Analysis**
- **Summary**

Wind Speed Measurements in Wind Energy

- **Wind Plant Operations**
 - **Power output validation**
 - **Control start-up and shut-down**
- **Wind Turbine Performance Evaluations**
 - **Power curve (wind turbine power output as a function of wind speed)**
- **Wind Energy Site Assessments**
 - **Wind speed distributions and corresponding turbine power output using power curves are used to determine the predicted Annual Energy Production, which is essential to Power Purchase Agreements**

Anemometer Calibration Protocol

Anemometer Output \Leftrightarrow Controlled Reference Speed




Rotation Rate
(i.e., Hz or rpm)



Voltage Signal
(Direct or
Conditioned)

Wind Generated from a
Controlled Wind Tunnel



Anemometers are primarily designed
to be linear instruments.

Perform a Least Squares Fit

\Rightarrow Linear Transfer Function

Anemometer Calibration Protocol

Common requirement from test standards is to perform anemometer calibration in a **CONTROLLED WIND TUNNEL**

- **ASTM D 5096-02**
“Standard Test Method for Determining Performance of a Cup Anemometer or Propeller Anemometer”- current edition 2002, first released in 1990
- **ASTM D 6011-96**
“Standard Test Method for Determining the Performance of a Sonic Anemometer/Thermometer” - current edition 2003, first released in 1996
- **ISO 17713-1**
“Meteorology - Wind measurements - Part 1: Wind tunnel test methods for rotating anemometer performance” - released in May 2007
- **ISO 16622**
“Meteorology – Sonic anemometers/thermometers – Acceptance test methods for mean wind measurements” - released in September 2002
- **IEC 61400-12-1**
“Power Performance Measurements of Electricity Producing Wind Turbines” - released in December 2005

Anemometer Calibration Protocol

Wind Tunnel Characteristic	Description	Minimum Requirements from Standards
Blockage	Ratio of the test anemometer frontal area over the test section cross-sectional area	IEC 61400-12-1: not to exceed 10% for open test sections and 5% for closed test sections ASTM & ISO: less than 5%, preferably closer to 1%
Flow Uniformity	Wind speed difference within the test section	IEC 61400-12-1: less than 0.2% ASTM & ISO: uniform velocity profile to within 1%
Horizontal Wind Gradient	Pressure difference over area covered by a rotating sensor	IEC 61400-12-1: must be less than 0.2%
Turbulence Intensity	Ratio of the wind speed standard deviation to its mean	IEC 61400-12-1: must be less than 2% ASTM & ISO: must be less than 1%
Air Density Uniformity	Air density difference within the test section	ASTM D 5096-02 & ISO 17713-1: less than 3% difference
Data Acquisition	Method of collecting reference wind speed data	IEC 61400-12-1: maintain resolution of 0.02 m/s at a sampling frequency of at least 1 Hz for 30 sec ASTM & ISO: maintain resolution of 0.02 m/s at a sampling frequency of at least 10 Hz for 30 to 100 sec
Wind Speed Capability	Controlled wind speed range	IEC 61400-12-1: generate controlled speeds from 4 to 16 m/s ASTM D 5096-02, ISO 17713-1: produce speeds at least from 0 to 50% of the test anemometer application range and maintained within +/- 0.2 m/s ASTM D 6011-96: generate speeds of at least 1 to 10 m/s and maintained within +/- 0.1 m/s or better ISO 16622: produce speeds over full application range of the test anemometer and maintained within +/- 0.2 m/s, preferably +/- 0.1 m/s

Anemometer Calibration Protocol

Sample Calibration Report

Description of Test Facility

Plot of Linear Regression

Photo of Test Setup

Report Signatory



630 Peña Drive, Suite 800
Davis, CA 95618
530.757.2264
www.otechwind.com

ANEMOMETER CALIBRATION REPORT

Test Date: 28 January 2010

Report Revision No: 0

Customer Information:
MEASNET (Contact: Assessment Team)
Ebertstrasse 96
26392 Wilhelmshaven
Germany



Model No: Windsensor (Riso) P2546A
Serial No: 6400
IUT Signal Power: 12 VDC with 10k Ω pull-up resistor
IUT Output: 0-10 VDC Waveform
Test Reference: Otech Calibration Campaign Rev 1

Wind Tunnel Test Facility
Otech Tunnel ID: WT2B
Type: Eiffel (open circuit, suction)
Test Section Size: 0.61 m x 0.61 m x 1.22 m
Manufacturer: Engineering Laboratory Design, Inc.

Data Acquisition
Hardware: National Instruments CDAQ-9172 USB 2.0 chassis with NI 9205 32-chan 16-bit AI module
Software: National Instruments LabVIEW 8.5
Signal Reduction Method for IUT: FFT to determine frequency

Measuring Equipment
Reference Speed: United Sensor Type PA Pitot-Static Tube sensed by an MKS Baratron Type 220D Differential Pressure Transducer (NIST traceable)
Amb. Pressure: Setra Model 270 Barometer (NIST Traceable)
Amb. Temperature: OMEGA HX94 SS RH Probe (NIST Traceable)
Relative Humidity: OMEGA HX94 SS RH Probe (NIST Traceable)

Test Conditions
Diff Pressure Transducer Position Correction = 1
Blockage Correction = 1
Mean Ambient Pressure = 101576 Pa
Mean Ambient Temperature = 22.7 deg C
Mean Relative Humidity = 41.2% RH
Mean Density = 1.1915 kg/cubic meter




Test Conditions

Plot of Residuals
 $\Delta V = V_{\text{measured}} - V_{\text{calculated}}$

Calibration Transfer Function and Regression Parameters:

$$V \text{ [m/s]} = 0.618 * f \text{ [Hz]} + 0.19$$

Slope = 0.618 m/s per Hz Correlation coeff. = 0.99999 Std. err. slope = 0.00097 m/s per Hz
Offset = 0.19 m/s Std. err. estimate = 0.021 m/s Std. err. offset = 0.01634 m/s



Note: Generic photo of test setup

Reference Speed [m/s]	Anemometer Output [Hz]	Speed Residual [m/s]	Ref. Speed Uncertainty
3.964	6.143	-0.020	0.514%
5.920	9.277	-0.001	0.488%
7.905	12.453	0.023	0.484%
9.885	15.683	0.007	0.475%
11.904	18.919	0.026	0.479%
13.864	22.129	0.003	0.475%
15.823	25.365	-0.038	0.471%
14.835	23.695	0.006	0.470%
12.882	20.571	-0.017	0.477%
10.874	17.278	0.010	0.473%
8.885	14.038	0.023	0.473%
6.915	10.875	0.007	0.495%
4.924	7.710	-0.029	0.488%

Approved by: Rachael Coquilla
Chief Engineer

Linear Transfer Function

Test Data with Speed Residuals and Uncertainty

This document reports that the above IUT was tested at Otech Engineering, Inc., a wind tunnel laboratory accredited in accordance with the recognised International Standard ISO/IEC 17025:2005 (Certificate number CL-126). This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communiqué dated January 2009). This report shall not be reproduced except in full, without written approval from Otech Engineering, Inc.




Anemometer Calibration Uncertainty

UNCERTAINTY



an estimated error range
for a measured variable



Bias, β
Systematic errors
typically Type B

Precision, σ
Random errors
typically Type A

**IMPORTANCE OF
UNCERTAINTY**

- 1) Defines the degree of error in a system
- 2) Identifies the errors in a system

Anemometer Calibration Uncertainty


IEC 61400-12-1

Anemometer Calibration Uncertainty is based on the errors in the calculated reference wind speed from the Pitot-static tube system of a controlled wind tunnel facility.

Controlled Wind Tunnel



Pitot-static Tube System



IEC 61400-12-1 Uncertainty Analysis

Reference Wind Tunnel Speed Equation using a Pitot-static tube system

$$V = k_b \sqrt{\frac{2k_c}{C_h} \frac{\Delta p}{\rho}}$$



Calculated Variables:

$$\rho = \frac{1}{T} \left[\frac{P}{R_{air}} - 0.01\phi P_w \left(\frac{1}{R_{air}} - \frac{1}{R_w} \right) \right]$$

$$P_w = 0.0000205 \exp(0.0631846T)$$

$$R_{air} = \frac{R}{M_{air}} \quad R_w = \frac{R}{M_w}$$

Measured Variables:

$$P, T, \phi, \Delta p, k_b, k_c$$

Defined Variables:

$$M_{air} \text{ and } M_w$$

IEC 61400-12-1 Uncertainty Analysis

Uncertainty in Wind Tunnel Reference Speed

Uncertainty, $U_V \Rightarrow f(k_b, k_c, C_h, R, P, T_K, \Delta p, \phi)$



$$U_V = \sqrt{B_V^2 + (tS_V)^2}$$

Bias or Systematic
Errors (Type B)

- Instrument manufacturer's spec
- Instrument calibration errors
- Data acquisition errors
- Correction factor errors

Precision or Random
Errors (Type A)

Standard deviations in
instrument readings

$\Delta p, P, T_K, \phi$

t-value at 95% confidence

$t = 1.96$ [Coleman & Steele (2009)]

IEC 61400-12-1 Uncertainty Analysis

Reference Speed Error Propagation



Propagation of Bias Errors:

$$B_V = \sqrt{\left(\frac{\delta V}{\delta k_b} B_{k_b}\right)^2 + \left(\frac{\delta V}{\delta k_c} B_{k_c}\right)^2 + \left(\frac{\delta V}{\delta C_h} B_{C_h}\right)^2 + \left(\frac{\delta V}{\delta R} B_R\right)^2 + \left(\frac{\delta V}{\delta P} B_P\right)^2 + \left(\frac{\delta V}{\delta T} B_T\right)^2 + \left(\frac{\delta V}{\delta \Delta p} B_{\Delta p}\right)^2 + \left(\frac{\delta V}{\delta \phi} B_\phi\right)^2}$$

Propagation of Precision Errors:

$$S_V = \sqrt{\left(\frac{\delta V}{\delta P} S_P\right)^2 + \left(\frac{\delta V}{\delta T} S_T\right)^2 + \left(\frac{\delta V}{\delta \Delta p} S_{\Delta p}\right)^2 + \left(\frac{\delta V}{\delta \phi} S_\phi\right)^2}$$

IEC 61400-12-1 Uncertainty Analysis

$$\left. \begin{aligned} \rho &= \frac{1}{T} \left[\frac{P}{R_{air}} - 0.01\phi P_w \left(\frac{1}{R_{air}} - \frac{1}{R_w} \right) \right] \\ P_w &= 0.0000205 \exp(0.0631846T) \\ R_{air} &= \frac{R}{M_{air}} \quad R_w = \frac{R}{M_w} \end{aligned} \right\} \begin{array}{l} \text{Original} \\ \text{Velocity Equation} \\ V = k_b \sqrt{\frac{2k_c}{C_h} \frac{\Delta p}{\rho}} \end{array}$$

Complete Velocity Equation



$$V = k_b \sqrt{\frac{2k_c \Delta p R T}{C_h [P M_{air} - 2.05 \times 10^{-7} \phi \exp(0.0631846T) (M_{air} - M_w)]}}$$

Derive to get the following sensitivity coefficients:

$$\frac{\delta V}{\delta k_b} \quad \frac{\delta V}{\delta k_c} \quad \frac{\delta V}{\delta C_h} \quad \frac{\delta V}{\delta R} \quad \frac{\delta V}{\delta P} \quad \frac{\delta V}{\delta T} \quad \frac{\delta V}{\delta \Delta p} \quad \frac{\delta V}{\delta \phi}$$


IEC 61400-12-1 Uncertainty Analysis

IEC 61400-12-1 Uncertainty Analysis for Otech Engineering Wind Tunnel Facilities




IEC 61400-12-1 Uncertainty Analysis

Rotating Anemometer Output Measurement in Otech Engineering Wind Tunnel Facilities



IEC 61400-12-1 Uncertainty Analysis

LabVIEW Data Acquisition System in Otech Engineering Wind Tunnel Facilities



**Anemometer
Output Signal**


**Reference
Wind Speed
Reading**

**Test
Conditions**

**Regression
Analysis**

IEC 61400-12-1 Uncertainty Analysis


Reference Wind Speed Uncertainty in Otech Engineering Wind Tunnel Facilities




Wind Speed (m/s)	Anemometer Output (Hz)	Wind Speed STD (m/s)	Wind Speed Bias Error (m/s)	Wind Speed Random Error (m/s)	Wind Speed Uncertainty (m/s)	Relative Wind Speed Uncertainty
3.99	6.25	0.003	0.019	0.003	0.020	0.51%
5.97	9.40	0.005	0.028	0.005	0.030	0.50%
7.96	12.59	0.004	0.037	0.004	0.038	0.48%
9.94	15.76	0.007	0.047	0.007	0.049	0.49%
11.92	18.96	0.007	0.056	0.007	0.057	0.48%
13.92	22.23	0.010	0.065	0.010	0.068	0.49%
15.91	25.56	0.011	0.075	0.011	0.078	0.49%
17.90	28.73	0.012	0.084	0.012	0.087	0.49%
19.90	32.02	0.011	0.093	0.011	0.096	0.48%
21.87	35.13	0.012	0.102	0.012	0.105	0.48%
23.87	38.42	0.014	0.112	0.014	0.115	0.48%
25.84	41.53	0.013	0.121	0.013	0.124	0.48%

$$U_V \approx 0.5\%$$

Expanded Uncertainty Analysis



Anemometer Under Test



$$U_{cal} = \sqrt{(U_V)^2 + (U_{IUT})^2 + (U_{LR})^2}$$

Reference
Wind Speed
Uncertainty

Anemometer
Output Uncertainty

Linear
Regression
Uncertainty


Expanded Uncertainty Analysis

Anemometer Output Uncertainty

$$U_{IUT} = \sqrt{B_{IUT}^2 + (tS_{IUT})^2}$$

Bias or Systematic
Errors (Type B)

Data acquisition errors
(FFT or pulse count method)



Precision or Random
Errors (Type A)

Standard deviations in
anemometer frequency or
voltage output
(FFT method)

t-value at 95% confidence
 $t = 1.96$ [Coleman & Steele (2009)]

Anemometer Output Uncertainty

$$U_{IUT} \cong 0.5\% - 1.5\%$$

Expanded Uncertainty Analysis

Linear Regression Uncertainty

Case 1

$$U_{LR_{Case1}} = \sqrt{\left(t \frac{STE_V}{V_i} \right)^2}$$

Standard Error of Estimate

$$STE_V = \sqrt{\frac{\sum_{i=1}^N (V_i - mf_i - b)^2}{N-2}}$$

Case 2

$$U_{LR_{Case2}} = \sqrt{\underbrace{\left(t \frac{STE_m}{V_i} f_i \right)^2}_{\text{Uncertainty in the Slope}} + \underbrace{\left(t \frac{STE_b}{V_i} \right)^2}_{\text{Uncertainty in the Offset}}}$$

Standard Error of Slope

Standard Error of Offset

$$STE_b = \sqrt{STE_V^2 \left(\frac{1}{N} + \frac{\bar{f}^2}{SS_f} \right)} \quad STE_m = \sqrt{\frac{STE_V^2}{SS_f}}$$

Sum of the Squares in Anemometer Output



$$SS_f = \sum_{i=1}^N f_i^2 - \frac{\left(\sum_{i=1}^N f_i \right)^2}{N}$$

Linear Regression Uncertainty (Case 1 or Case 2)

$$U_{LR} \cong 0.2\% - 5.0\%$$

Expanded Uncertainty Analysis

Sample Expanded Uncertainty Analysis



$STE_V = 0.031 \text{ m/s}$

Wind Speed (m/s)	Wind Speed Uncertainty	Anem. Output (Hz)	Anem. STD (Hz)	Anemometer DAQ Bias Error (Hz)	Anemometer Random Error (Hz)	Anemometer Output Uncertainty	Case 1 Regression Uncertainty	Case 2 Regression Uncertainty	Anemometer Calibration Uncertainty
3.99	0.51%	6.25	0.061	0.00063	0.061	1.94%	1.52%	1.07%	2.52%
5.97	0.50%	9.40	0.082	0.00094	0.082	1.74%	1.02%	0.74%	2.08%
7.96	0.48%	12.59	0.119	0.00126	0.119	1.90%	0.76%	0.58%	2.10%
9.94	0.49%	15.76	0.111	0.00158	0.111	1.41%	0.61%	0.49%	1.62%
11.92	0.48%	18.96	0.116	0.00190	0.116	1.22%	0.51%	0.43%	1.41%
13.92	0.49%	22.23	0.130	0.00222	0.130	1.17%	0.44%	0.39%	1.34%
15.91	0.49%	25.56	0.145	0.00256	0.145	1.13%	0.38%	0.36%	1.29%
17.90	0.49%	28.73	0.169	0.00287	0.169	1.18%	0.34%	0.34%	1.32%
19.90	0.48%	32.02	0.172	0.00320	0.172	1.08%	0.30%	0.33%	1.22%
21.87	0.48%	35.13	0.172	0.00351	0.172	0.98%	0.28%	0.32%	1.13%
23.87	0.48%	38.42	0.174	0.00384	0.174	0.91%	0.25%	0.31%	1.06%
25.84	0.48%	41.53	0.208	0.00415	0.208	1.00%	0.23%	0.30%	1.13%

$U_V \approx 0.5\%$

$U_{IUT} \approx 1.3\%$

$U_{LR} \approx 0.5\%$

$U_{IUT} \approx 1.5\%$

Summary

- **Wind Speed Measurements in Wind Energy**
 - **Wind Plant Operations**
 - **Wind Turbine Performance Evaluations**
 - **Wind Energy Site Assessments**
- **Anemometer Calibration Protocol**
 - **ASTM D 5096-02, ASTM D 6011-96, ISO 17713-1, ISO 16622**
 - **IEC 61400-12-1 (2005) “Power Performance Measurements of Electricity Producing Wind Turbines”**

Summary

- **Anemometer Calibration Uncertainty**
 - **IEC 61400-12-1 Uncertainty Analysis**

$$U_V = \sqrt{B_V^2 + (tS_V)^2} \longrightarrow U_V \cong 0.5\%$$

- **Expanded Uncertainty Analysis**

$$U_{IUT} = \sqrt{B_{IUT}^2 + (tS_{IUT})^2} \longrightarrow U_{IUT} \cong 0.5\% - 1.5\%$$

$$\left. \begin{aligned} U_{LR_{Case1}} &= \sqrt{\left(t \frac{STE_V}{V_i}\right)^2} \\ U_{LR_{Case2}} &= \sqrt{\left(t \frac{STE_m}{V_i} f_i\right)^2 + \left(t \frac{STE_b}{V_i}\right)^2} \end{aligned} \right\} U_{LR} \cong 0.2\% - 5.0\%$$

$$U_{cal} = \sqrt{(U_V)^2 + (U_{IUT})^2 + (U_{LR})^2} \longrightarrow U_{cal} \cong 0.7\% - 5.2\%$$