

# ***Considerations in Anemometer Calibration***

Presented by:  
**John Obermeier, President**  
Otech Engineering, Inc.

In collaboration with:  
**Rachael Coquilla, Ph.D. candidate**  
Dept. Mechanical & Aeronautical Engineering  
University of California at Davis

# Outline

- ✚ **Current Standards and Test Protocols**
- ✚ **Anemometer Calibration**
- ✚ **Quality Control**
- ✚ **Calibration Comparisons**
- ✚ **Calibration Standard Considerations**
- ✚ **Summary**

## Current Standards and Test Protocols

### *Current Test Standards for cup and propeller anemometers*

- **ASTM D5096-02**
- **ISO 17713**

starting threshold  
distance constant  
transfer function \*\*  
off-axis response

### **Non-standard Protocols**

- **Measnet**
- **Otech Engineering, Inc.**
  - ...moving vehicle method
  - ...wind tunnel method



# Anemometer Calibration

## **Background**

- 30 year history
- Most widely used instrument in wind industry (over 150,000+ units installed)
- Most tested anemometer in the world

## **Performance characteristics**

- Rugged and survives high wind environments
- Linear range matches wind turbine energy producing speed range
- Independent of temperature
- Maintains performance for extensive periods



**NRG #40 Cup Anemometer**



# Anemometer Calibration

## Generation of an anemometer transfer function


- ✚ Steady state wind speeds at set intervals
- ✚ Increasing and decreasing speeds
- ✚ Include specified speed range
- ✚ Produce slope and offset from linear transfer function
- ✚ Provide a measure of uncertainty



# Anemometer Calibration

Uncertainty in Velocity Measurement

$$U_V = \sqrt{(B_V)^2 + (tS_V)^2}$$



# Quality Control


Recall sample transfer function

$$U = 0.7707 f + 0.318$$

Converted transfer function  
using forced intercept


$$U = 0.7693 f + 0.35$$

“slope+k” value





# Quality Control

Histogram of Slope+K Values from OEI Vehicle Tests  
Dec 2002 – Jul 2005 (11,834 new NRG #40 Cup Anemometers)  
Standard Deviation = 0.0031 m/s per Hz  
95 % of the distribution are included within  $\pm 0.8$  % of the mean




# Calibration Comparisons

Comparison of five NRG #40 anemometer calibrations (year 2002) between the OEI Vehicle and CRES Wind Tunnel test methods




# Calibration Comparisons

Comparison of 100 NRG #40 anemometer calibrations (year 2005) between the OEI Vehicle and OEI Wind Tunnel test methods



◆ OEI Vehicle    ■ OEI Wind Tunnel    ● Consensus Value



— OEI Vehicle    — OEI Wind Tunnel


## Calibration Standard Considerations

- ✚ Must agree upon a consensus standard anemometer calibration procedure and report process (i.e. Measnet).
- ✚ Consensus must remain an open process.
- ✚ Maintain a relation to a standard instrument via round robin comparisons (i.e. RR-3 experiment).

Test Protocol	Range (m/s)	# Points	Dwell Time (sec)	Hysterises Check	Interval Duration (sec)
ASTM D 5096-02	0 - 50% FS	10	300 - 1000	yes	30 - 100
OEI Vehicle Test	4.5 - 25	10	396	yes	18 - 98
OEI Wind Tunnel	4 - 26	12	360	yes	30
Measnet	4 - 16	13	390	yes	30

# Calibration Standard Considerations

Power, Speed Frequency, and Energy Distribution for a range of wind speeds (using GE 1.5 MW wind turbine power curve)



# Calibration Standard Considerations

## Considerations for advanced calibrations:

### ✦ **Over-speeding**

*Definition: a cup anemometer's over-estimation of the mean wind due to its faster dynamic response at the initial exposure to the wind than at the decrease of the same wind.*

### ✦ **Off-axis**

**ASTM D5096-02 recommended testing**

- **test range  $\pm 30^\circ$  off-axis**
- **increments of  $5^\circ$  angles**

### ✦ **Temperature dependence**

## Summary

- ✚ To add value in the manufacture of an anemometer, it is recommended to maintain quality control by using anemometer calibration and a statistical guideline (“Slope+K”). This assures buyers that the product clearly conforms to a defined performance criteria.
- ✚ When a reliable calibration transfer function is available, it is recommended that users apply the tested transfer function rather than a generic transfer function.
- ✚ Uncertainty in reference wind speeds for anemometer calibration are in the range of 2%. Uncertainty claims lower than NIST using LDA technology should be questioned.
- ✚ An anemometer calibration standard needs to be defined.

## **Related Measurement Standards and References**

- # **ASTM D5096-02 Standard Test Method for Determining the Performance of a Cup Anemometer and Propeller Anemometer**
- # **ASTM D4430-00e1 Standard Practice for Determining the Operational Comparability of Meteorological Measurements**
- # **ASTM D6011-96(2003) Standard Test Method for Determining the Performance of a Sonic Anemometer/Thermometer**
- # **ASTM D7145-05 Standard Guide for Measurement of Atmospheric Wind and Turbulence Profiles by Acoustic Means**
- # **ASTM D5741-96(2002)e1 Standard Practice for Characterizing Surface Wind Using a Wind Vane and Rotating Anemometer**
- # **ASTM D5527-00(2002)e1 Standard Practices for Measuring Surface Wind and Temperature by Acoustic Means**
- # **“The Maximum Type 40 Anemometer Calibration Project” Thomas J. Lockhart, CCM. Cnet, April 1998**